



DROPPING BALL APPARATUS

Standard: BS 4551-1, 6463-4

Sand Equivalent Value Test

Use to measure the consistency of cement mortars, this instrument allows a 25 mm diameter acrylic ball fall freely from a standard height.

Technical Specifications :

Model Number	CL 3002 X / 001
Height of Drop	250 mm
Approx. Weight	8 kg

Set Consists Of :

Model Number	Parts Description	Qty
CL 3002 X / 001 – P 001	Dropping Ball Device Stand	1set
CL 3002 X / 001 – P 002	Acrylic Ball	1no.
CL 3002 X / 001 – P 003	Mortar Ring Mould Ø 100 x 25 mm	1no.
CL 3002 X / 001 – P 004	Penetration Measuring Device	1no.
CL 3002 X / 001 – P 005	Rubber Suction Hand Pump	1no.
CL 7020 X / 001	Dial Indicator 25 x 0.01 mm	1no.

*1 Copy of Manual Instruction