



GILLMORE APPARATUS

Standard: ASTM C91, C141, C266 / AASHTO T154

Determine of the Setting Time Cement

Used to determine the setting time of cement. The apparatus consist of two horizontal arms, which carry two weighted steel needles, which are precision machined to meet the specified requirements.

Technical Specifications :

Model Number	CL 3003 X / 001
Initial Setting Needle Diameter	2.12 mm
Final Setting Needle Diameter	1.06 mm
Initial Setting Weight Piece	113 Gram
Final Setting Weight Piece	453.6 Gram
Dimension (mm)	200 (L) x 100 (W) x 200 (H)
Approx. Weight	2.1 kg

*1 Copy Of Manual Instructions