



MOISTURE CONDITION VALUE (MCV)

Standard: BS 1377:4, EN 13286-46

This apparatus is used to measure the minimum comparative effort required to produce near full compaction of a soil, and the rate at which a sample of chalk lumps are crushed. The unit can be used to classify chalk as a fill material with saturated moisture content. The apparatus comprises a robust frame where a rammer falls, mould, scale, counter, accessories.

Technical Specifications :

Model Number	CL 5055 X / 001
Approx. Weight	60 kg

Set Consists Of :

Model Number	Specifications	Qty
CL 5055 X / 001 - P001	Moisture Condition Mould	1no.